



# Piccolo Basic - Plus - Speech

PC CONTROLLED AUDIOMETER



## DESCRIPTION

Piccolo is an audiometer controllable through any Windows computer.

Piccolo is available in three different models:

- Basic*: only PTA – AC
- Plus*: only PTA – AC / BC
- Speech*: PTA & Speech – AC / BC

When controlled through the computer using Inventis Maestro software, Piccolo is connected to it via USB.

## CLASSIFICATION

- Basic: IEC 60645-1 / ANSI S3.6: Type 4
- Plus: IEC 60645-1 / ANSI S3.6: Type 3
- Speech: IEC 60645-1 / ANSI S3.6: Type 3B

## AVAILABLE SIGNALS

- Stimulus: pure tone, warble tone
- Masking: NBN, WN, SN (Piccolo Speech only)

Only Piccolo Speech:

- 2 external inputs for speech audiometry
- MIC input for live speech audiometry

## SIGNALS SPECIFICATION

- Attenuator step: 1 dB, 3 dB, 5 dB
- Presentation: Continuous, Pulsed (0.5, 1 and 2 Hz)
- Warble: 5 Hz sine wave modulating signal

## AVAILABLE OUTPUTS AND TRANSDUCERS

- AC: TDH-39 / DD45 / DD65 headphones, ER-3 / IP30 insert earphones
- BC: B-71 bone vibrator (Piccolo Plus / Speech only)

## AVAILABLE TESTS

- Pure Tone audiometry
- Automatic Pure Tone audiometry
- Speech audiometry (Piccolo Speech only)
- QuickSIN™ (optional for Piccolo Speech only)
- Master Hearing Aid (Piccolo Speech only)

## PURE TONE: FREQUENCIES AND MAXIMUM LEVELS (dB HL)\*

Freq. (Hz)	AC TDH-39 DD45	AC DD65	AC ER-3 IP30	AC ER-5**	BC
125	80	80	90	90	-
250	100	95	105	100	45
500	110	110	110	110	65
750	115	110	115	120	70
1.000	120	110	120	120	75
1.500	120	110	120	120	80
2.000	120	110	120	115	80
3.000	120	110	120	115	75
4.000	120	110	110	110	75
6.000	105	95	95	100	55
8.000	95	90	90	90	50

(\* ) Obtainable only if the audiometer is powered with the mains adapter

(\*\* ) ER-5 no more available for purchasing but supported

## SPEECH AUDIOMETRY: MAXIMUM LEVELS (dB HL)\*

AC TDH-39 DD45	AC DD65	AC ER-3 IP30	AC ER-5**	BC
100	90	100	100	55

(\* ) Only available for the Speech version.

(\*\* ) ER-5 no more available for purchasing but supported

## PATIENT – OPERATOR COMMUNICATION

- Talk over: external microphone included
- Patient response trigger

## CALIBRATION

- Validity: 12 months
- All the parameters set through the device software

## COMPUTER INTERFACE

- Connection: USB (driverless)
- Compatible software: Inventis Maestro
- Minimum requirements: refer to Maestro software specifications

## POWER SUPPLY

- Maximum consumption: 8 Watts
- Power supply: 6V, 1.67A cont., through an external medical grade 100-240 Vac 50/60 Hz power supply

## MECHANICS

- Size (WxDxH): 16 x 16 x 3 cm / 6.3 x 6.3 x 1.2 in
- Weight: 300g / 10.6 oz

## FREIGHT PACKING

- Size (WxDxH): 40 x 30 x 18 cm / 15.8 x 11.8 x 7.1 in
- Gross weight: 2.7 kg / 5.9 lbs

## APPLICABLE STANDARDS

- Calibration: ISO 389-1, ISO 389-2, ISO 389-3
- Electrical safety: IEC 60601-1, Class I type B
- EMC: IEC 60601-1-2

## CE CERTIFICATE

- MDR 2017/745/EU Classification: Class IIa
- Classification rule (Annex VIII, 2017/745): 10
- Notified body: TÜV SÜD Product Service GmbH (0123)

**PRODUCT CODES**

Code 10327 - Piccolo model Basic  
Code 10329 - Piccolo model Plus  
Code 10331 - Piccolo model Speech

**INCLUDED PARTS**

- DD45, TDH-39 or DD65 headphones
- B71 bone vibrator (Piccolo Plus / Speech only)
- Patient response switch
- Clip-on microphone for talk-over
- Medical grade power supply
- USB connection cable
- Jack 3.5mm connection cable (Piccolo Speech only)
- Carrying bag
- User manual
- Inventis Software Suite

**OPTIONAL PARTS (with order code)**

- 10833: ER-3C insert earphones
- 11748: IP30 insert earphones
- 10179: Amplivox Audiocups noise excluding enclosures for TDH-39 / DD45
- 10533: QuickSIN™ test license

Piccolo is developed by:

**INVENTIS S.r.l.**

CORSO STATI UNITI, 1/3  
35127 PADOVA – ITALY  
PHONE: +39.049.8962 844  
FAX: +39.049.8966 343  
info@inventis.it  
www.inventis.it

Follow us on LinkedIn

<https://www.linkedin.com/company/inventis-srl>

The Inventis Quality System complies with ISO 13485 standard.

Inventis® is a registered trademark of INVENTIS S.r.l.