



Maestro

SOFTWARE

DESCRIPTION

Maestro is a software for managing patient data that can interface with all the Inventis device, controlling or getting exams results from them.

OPERATING MODES

- Noah module
- OtoAccess module
- Stand-Alone with proprietary database
- Stand-Alone without database functions (file based)
- Third-party Applications (exchange data with external applications): MII proprietary protocol and GDT standard

AVAILABLE TESTS

- Audiology exams
- REM
- HIT
- Evoked Potentials
- Video-Otoscopy
- Video and Images acquisition
- Questionnaire
- Session Data

COMPATIBLE DEVICES

- Triangle (Audiometer)
- Bell (Audiometer)
- Harp (Audiometer)
- Piano (Audiometer)
- Piccolo (Audiometer)
- Cello (Audiometer)
- Trumpet (REM, Audiometer)
- Drum (HIT)
- Flute (Middle-Ear Analyzer)
- Clarinet (Middle-Ear Analyzer)
- Viola (Middle-Ear Analyzer, Audiometer)
- Timpani (Handheld Tympanometer, Audiometer)
- Celesta (Evoked Potentials system)
- Delfino (Otoscope)
- Harmonica (Otoscope)
- MediCam (Video-endoscopy camera)
- MediCam HD (HD Video-endoscopy camera)
- Flexia 2 (Video-Nasopharyngoscope)
- Analog-to-USB video converter
- Analog-to-USB video converter HD

COMPATIBLE ACCESSORIES

- Maestro Control Panel

INTERFACE WITH EQUIPMENT

USB connection

PC MINIMUM REQUIREMENTS

CPU: Intel® i5, 6th generation or above, 8 GB RAM
 Operating System: Windows 10 and Windows 11, 64 bits
 Display resolution: suggested 1920 x 1080 or higher (zoom factor 100% or 125%)
 Local network (only for db installation in network folder): 1 Gb/s
 Free space on HD: 3 GB

NOAH MINIMUM REQUIREMENTS

NOAH 4.4 or greater (or NOAH Engines supporting NOAH 4.4 or greater)

OTOACCESS MINIMUM REQUIREMENTS

OtoAccess2 version 2.4.1.18 or greater

SOFTWARE SUPPORT

Remote assistance tool: included

SOFTWARE UPDATES

Available for free

MII INTERFACE

Library for interfacing Maestro with 3rd party software
 Available upon request (IT technician intervention requested)

APPLICABLE STANDARDS

Software life cycle: IEC 62304
 Health software: IEC 82304-1, IEC 81001-5-1

CE CERTIFICATE

MDR 2017/745/EU Classification: Class IIa
 Classification rule (Annex VIII, 2017/745): 11
 Notified body: TÜV SÜD Product Service GmbH (0123)

Maestro is developed by:

INVENTIS S.r.l.

CORSO STATI UNITI, 1/3
 35127 PADOVA – ITALY
 PHONE: +39.049.8962 844
 FAX: +39.049.8966 343
 info@inventis.it
 www.inventis.it

Follow us on LinkedIn

<https://www.linkedin.com/company/inventis-srl>

The Inventis Quality System complies with ISO 13485 standard.

Inventis® is a registered trademark of INVENTIS S.r.l.